

Culture Media

ENDO-050P

Tech Sheet

Use

For enumerating total coliforms by the membrane filtration method.

Formula

Yeast Extract	1,5 g/l
Casitone	5 g/l
Thiopeptone	5 g/l
Tryptose	10 g/l
Lactose	12,5 g/l
Sodium Desoxycholate	0,1 g/l
Dipotassium Phosphate	4,38 g/l
Monopotassium Phosphate	1,38 g/l
Sodium Chloride	5 g/l
Sodium Lauryl Sulfate	0,05 g/l
Sodium Sulfite	2,1 g/l
Basic Fuchsin	1,05 g/l

Precautions

- For industrial use only .
- Follow proper procedures for handling and disposing of infectious materials laboratory .

Directions

- 1) Remove the protective bag, remove the plug from the base .
- 2) Place the monitor on the connector previously placed on the manifold or drilled silicone stopper.
- 3) Remove the lid of the funnel and pour the sample into the funnel.
- 4) Apply vacuum just long enough to pull the sample through the filter.
- 5) Remove the rubber stopper of the collector to remove the remaining empty and then replace the cap on the collector.
- 6) Remove the funnel and distribute the culture media over the membrane filter. Be careful not to touch the filter with the tip of the ampoule.
- 7) Apply vacuum briefly to allow the culture media to go through the membrane and be absorbed by the cellulose pad, the time required for the medium crop does not fully rest on the filter (this must not be supersaturated or dry) and flow to the pad on the bottom of the monitor.
- 8) Remove the lid from the funnel and cover the base with it. Remove the base from the manifold or silicone stopper and close the bottom port of the base with the previously removed plug.
- 9) Incubate the so created Petri dish, inverted according to the culture media.



Reading results

Incubation: For 18 to 24 hours at 35 ± 2°C

Interpretation of results: All colonies that are red and have the characteristic metallic sheen are considered coliforms.

Limitations

1) Since the nutritional requirements of organisms vary, some strains may be encountered that fail to grow or grow poorly on this medium.

QC Test

QC TEST MICROBIOLOGICAL

Cultural characteristics observed after 24 hs at 35 ± 2°C

MICROORGANISMS (ATCC)	RECOVERY RATE
<i>Escherichia coli</i> (25922)	≥ 70%
<i>Staphylococcus aureus</i> (6538)	Inhibition

Cultural characteristics observed after 72 hs at 37°C

STERILITY	RECOVERY RATE
	No growth

pH: 7.2 ± 0.2

Shelf life

12 months.

Storage

Store at room temperature.

Code

ENDO-050

Packaging

Box of 50 glass ampoules .

For industrial use only.